

Technical Data Sheet

Doc No – Ab/TDS/CW-12

Rev No : 01

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Universal Centrifuge Tube Rack, PP

Valid for Catalog No. P20212

1. Description/Specification

1.1 Description:

Reusable Polypropylene (PP) Universal Centrifuge Tube Rack.

1.2 Design:

Rectangular Design with holding a wide range of tubes and vessel.

1.3 Performance Specification:

1.3.1 Capacity : 15 ml and 50 ml Centrifuge tube.

1.3.2 Universal Fit for 15 ml and 50 ml Centrifuge Tube Respectively.

1.3.3 Places: 12 (4x3 Array)

1.4 Material:

1.4.1 Polypropylene (PP)

1.4.2 Comply with requirement of Directive 2002/72/EC and requirement of FDA Regulation 21 CFR 177.1520. olefins polymers.

1.4.3 Complies with the requirement of Indian Standard IS: 10910 on “ Specification for Polypropylene and its copolymer for its safe use in contact with food stuffs, Pharmaceutical and Drinking water”

1.4.4 Free from heavy metals according to ELV directive 2000/53/EC and its following amendments.

1.5 Color:

Box : Blue

1.6 Sterilization:

Can be autoclaved at 15 psi, 121⁰C for 20 minutes



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttarakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslaware.com
W : www.abdoslaware.com

Technical Data Sheet

1.7 Quality control:

Testing of various characteristic in accordance with the valid Specification

1.8 other Information:

Reusable, can be autoclaved up to 10 autoclave cycle.

1.9. Basic Features

1.9.1 Made up of High pure virgin Polypropylene (PP)


1.9.1 Temperature range: -90⁰C to +121⁰C


1.9.3 Autoclavability: Yes, Autoclavable at 15 psi and 121⁰C for 20 minutes

1.9.4 Chemical resistance: See website.

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121⁰C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50⁰C before removing product.


Quality Assurance Incharge
Abdos Labtech Pvt. Ltd.



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttrakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslabware.com
W : www.abdoslabware.com