

Technical Data Sheet

Doc No – Ab/TDS/GLP-131

Rev No : 03

TECHNICAL DATA SHEET

Abdos Labtech Pvt. Ltd.

Measuring Cylinder – PP – Class B

Valid for Catalog No. P50301, P50302, P50303, P50304, P50305, P50306, P50307, P50308

1. Description/Specification

1.1 Description:

PP Measuring Cylinder

1.2 Design:

1.2.1 Body : Cylindrical with generous Pour spout for smooth of liquid While dispensing

1.2.2 Base : Hexagonal Base for stability

1.3 Performance Specifications:

1.3.1 Volume : **P50301** (10 ml), **P50302** (25 ml), **P50303** (50 ml), **P50304** (100 ml), **P 50305** (250 ml),

P50306 (500ml), **P50307** (1000ml), **P50308** (2000ml)

1.4 Material:

1.4.1 PP- Polypropylene (PP)

1.4.2 Comply with requirement of Directive 2002/72/EC and requirement of FDA Regulation 21 CFR 177.1520. olefins polymers.

1.4.3 Complies with the requirement of Indian Standard IS: 10910 on “ Specification for Polypropylene and its copolymer for its safe use in contact with food stuffs, Pharmaceutical and Drinking water”

1.4.4 Free from heavy metals according to ELV directive 2000/53/EC and its following amendments.

1.5 Color:

Translucent (Natural)

1.6 Graduations:

Molded in Large and easy to read graduation

Meet accuracy requirements of class B under “DIN 12681/ISO6706”



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttarakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslabware.com
W : www.abdoslabware.com

Technical Data Sheet

1.7 Sterilization:

Can be autoclaved at 15 Psi, 121 °C for 20 minutes

1.8 Quality control:

Testing of various characteristic in accordance with the valid Specification

1.9 other Information:

Reusable, can be autoclaved up to 10 autoclave cycle.

Pack Size: 01 pieces/case

2. Basic Features

2.1 Made up of High purity Virgin Polypropylene (PP)


2.1 Temperature range: -30°C to +121°C

2.3 Autoclavability: Yes, Autoclavable at 15 psi and 121°C for 20 minutes

2.4 Chemical resistance: See website.

Autoclaving Guideline for Plastics

All plastic consumables are autoclavable at 121°C for 20 minutes, 15 psi/1atm. Do not use a dry cycle as this may cause deformation or weakening of the plastic. Allow temperature within the autoclave to return to at least 50°C before removing product.


Quality Assurance Incharge
Abdos Labtech Pvt. Ltd.



Shiv Ganga Industrial Estate,
Plot no. 1 Lakeshwari, Bhagwanpur,
Roorkee - 247667 Uttrakhand, India
An ISO 9001:2008 Company
A Division of Abdos

Regd. Office :
13A, Government Place (East)
Kolkata – 700069, India
CIN : U85100WB2008PTC123903

P : +91 (0) 1332 235277
E : info@abdoslabware.com
W : www.abdoslabware.com